



FLORIDA DEPARTMENT OF
Environmental Protection

Bob Martinez Center
2600 Blair Stone Road
Tallahassee, Florida 32399-2400

Ron DeSantis
Governor

Jeanette Nuñez
Lt. Governor

Noah Valenstein
Secretary

Electronic Submission

Air General Permit Re-Registration

You have successfully submitted an Air General Permit Registration to operate a **Ethylene Oxide Sterilizers** Facility under the authority of **Rule 62-210.310, F.A.C.**. Your registration was received on **July 03, 2019**. Unless you are notified by the Department of ineligibility to use the air general permit, you may use the air general permit thirty (30) days after giving notice to the Department.

Below is a copy of the details of your registration for your records.

Facility Information

Facility ID: 0251378
Facility or Business Name: MEDTRONIC INC.
Site Name: MIAMI LAKES SITE
Address Line 1: 14440 NW 60TH AVENUE
Address Line 2:
City/State/Zip Code: MIAMI LAKES, FL 33014

Mailing Address

Address Line 1: 14440 NW 60TH AVENUE
Address Line 2:
City/State/Zip Code: MIAMI LAKES, FL 33014

Owner

Name: Medtronic Inc.
Address Line 1: 14420 NW 60th Ave
Address Line 2:
City/State/Zip Code: Miami Lakes, FL 33014 2807

Facility Contact(s)

Name: MARK AESCHLEMAN
Address Line 1: 14420 NW 60th Ave
Address Line 2:
City/State/Zip Code: Miami Lakes, FL 33014 2807
Phone Number: (305) 364-1493
Extension:
Cell Number: (305) 968-7352
Fax Number: (305) 364-1525
E-mail Address: MARK.AESCHLEMAN@MEDTRONIC.COM

Correspondence Contact(s)

Name: JORGE BAHRI
Address Line 1: 14420 NW 60th Ave
Address Line 2:
City/State/Zip Code: Miami Lakes, FL 33014 2807
Phone Number: (305) 364-1493
Extension:
Cell Number: (786) 280-7790
Fax Number: (305) 364-1525
E-mail Address: jbahri@heartware.com

Notification Submitted By

Name: Mark Aeschleman
Phone Number: (305) 364-1493
E-mail Address: mark.aeschleman@medtronic.com

All information submitted was certified true, accurate, and correct to the best of the knowledge of the person whose name appears above.

If you have any questions or concerns about the information contained in this report, please contact the Small Business Environmental Assistance Program at (800) 722-7457 or by e-mail at Small.Business@floridadep.gov.

Facility Information¹**Ethylene Oxide Sterilizer Details:**

Vent Type	Date Initially Purchased from	Status	Control Device	Control	Date
HearWare_Miami Lakes_PermitApplic_NonCBI - 000002					

	Manufacturer		Required	Device	Installed
Sterilization Chamber	06/01/2018	New	No		

Facility Description:

This sterilizer is used to sterilize returned (used) heart pumps prior to inspection. The chamber size is 15'H x 17'W x 32.5'D with a volume of 4.8 cubic ft. The sterilization process uses a single, 100g, cartridge and the maximum concentration of EtO, in the chamber, is 735mg/L.

Material Usage:

Actual Usage (past 12 months): 0024

Highest Usage: From 01/01/2018 to 12/31/2018 was 0024

¹Items appearing in **RED** indicate a value or response that may cause your Air General Permit to be out of compliance.